

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081424\
 Data File : VN083313.D
 Acq On : 14 Aug 2024 20:07
 Operator : JC\MD
 Sample : PB162684TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162684TB

Quant Time: Aug 14 23:26:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	128089	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	254196	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	274194	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	123345	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	105493	57.861	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.720%
35) Dibromofluoromethane	8.171	113	84338	53.155	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.320%
50) Toluene-d8	10.565	98	319511	53.985	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.960%
62) 4-Bromofluorobenzene	12.847	95	133649	57.922	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	115.840%
Target Compounds						
					Qvalue	
16) Acetone	4.441	43	23627	33.118	ug/l	99
20) Methylene Chloride	5.283	84	12752	7.767	ug/l	84
25) 2-Butanone	7.488	43	16928	15.453	ug/l	92
43) Isopropyl Acetate	8.688	43	153377	31.921	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081424\
Data File : VN083313.D
Acq On : 14 Aug 2024 20:07
Operator : JC\MD
Sample : PB162684TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162684TB

Quant Time: Aug 14 23:26:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

