

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081424\
 Data File : VN083316.D
 Acq On : 14 Aug 2024 21:19
 Operator : JC\MD
 Sample : P3609-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 915-J-WS-081424

Quant Time: Aug 14 23:27:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	123251	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	241969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	252623	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	111058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	102801	58.598	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.200%
35) Dibromofluoromethane	8.171	113	81558	54.001	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.000%
50) Toluene-d8	10.565	98	302168	53.634	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.260%
62) 4-Bromofluorobenzene	12.847	95	125840	57.294	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	114.580%
Target Compounds						
16) Acetone	4.453	43	2556	3.723	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081424\
Data File : VN083316.D
Acq On : 14 Aug 2024 21:19
Operator : JC\MD
Sample : P3609-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
915-J-WS-081424

Quant Time: Aug 14 23:27:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

