

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087554.D
 Acq On : 14 Aug 2025 10:32
 Operator : JC\MD
 Sample : VN0814WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0814WBL01

Quant Time: Aug 14 15:23:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	284398	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	560321	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	503771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	234765	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	255137	52.871	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.740%
35) Dibromofluoromethane	8.153	113	170840	44.201	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.400%
50) Toluene-d8	10.547	98	608969	44.169	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.340%
62) 4-Bromofluorobenzene	12.829	95	222642	43.709	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087554.D
Acq On : 14 Aug 2025 10:32
Operator : JC\MD
Sample : VN0814WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBL01

Quant Time: Aug 14 15:23:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

