

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087557.D
 Acq On : 14 Aug 2025 12:03
 Operator : JC\MD
 Sample : Q2832-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TG-S01

Quant Time: Aug 14 15:24:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	245148	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	526488	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	494450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	224457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	226430	54.435	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.880%
35) Dibromofluoromethane	8.153	113	161920	44.585	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.180%
50) Toluene-d8	10.547	98	604289	46.646	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.300%
62) 4-Bromofluorobenzene	12.829	95	216028	45.136	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.280%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	60663	31.104	ug/l	100
18) Methyl Acetate	5.018	43	50482	10.302	ug/l	99
20) Methylene Chloride	5.277	84	19365	5.301	ug/l	94
43) Isopropyl Acetate	8.677	43	21470	1.996	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087557.D
Acq On : 14 Aug 2025 12:03
Operator : JC\MD
Sample : Q2832-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TG-S01

Quant Time: Aug 14 15:24:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

