

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087563.D
 Acq On : 14 Aug 2025 14:14
 Operator : JC\MD
 Sample : Q2838-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-11

Quant Time: Aug 14 15:25:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	224842	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	507918	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	489215	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	221171	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	218641	57.310	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.620%
35) Dibromofluoromethane	8.153	113	160097	45.695	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.380%
50) Toluene-d8	10.547	98	587894	47.040	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.080%
62) 4-Bromofluorobenzene	12.829	95	209709	45.418	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.840%
Target Compounds						
						Qvalue
16) Acetone	4.424	43	65693	36.725	ug/l	95
18) Methyl Acetate	5.018	43	46294	10.301	ug/l	98
20) Methylene Chloride	5.271	84	23296	7.155	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087563.D
Acq On : 14 Aug 2025 14:14
Operator : JC\MD
Sample : Q2838-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-11

Quant Time: Aug 14 15:25:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

