

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081518\
 Data File : VN050644.D
 Acq On : 15 Aug 2018 10:42
 Operator : MD\SY
 Sample : J4469-01DL 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 956-IW-13(19)DL

Quant Time: Aug 16 13:38:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	640130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	986945	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	872121	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	313757	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	412125	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
35) Dibromofluoromethane	7.59	113	387357	49.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.32%	
50) Toluene-d8	10.09	98	1411382	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
62) 4-Bromofluorobenzene	12.40	95	387303	39.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.06%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.83	96	48736	5.81	ug/l	89
44) Trichloroethene	8.84	130	79938	8.93	ug/l	99
64) Tetrachloroethene	10.63	164	364519	45.01	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081518\
Data File : VN050644.D
Acq On : 15 Aug 2018 10:42
Operator : MD\SY
Sample : J4469-01DL 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
956-IW-13(19)DL

Quant Time: Aug 16 13:38:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

