

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081518\
 Data File : VN050652.D
 Acq On : 15 Aug 2018 13:59
 Operator : MD\SY
 Sample : J4465-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 941-MW-02(23.8)

Quant Time: Aug 16 03:19:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	632097	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	957998	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	846547	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	319867	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	408899	51.32	ug/l	0.00
Spiked Amount						
			Recovery	=	102.64%	
35) Dibromofluoromethane	7.59	113	384814	50.31	ug/l	0.00
Spiked Amount						
			Recovery	=	100.62%	
50) Toluene-d8	10.09	98	1363861	47.38	ug/l	0.00
Spiked Amount						
			Recovery	=	94.76%	
62) 4-Bromofluorobenzene	12.40	95	390341	41.05	ug/l	0.00
Spiked Amount						
			Recovery	=	82.10%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	2.18	62	9882	1.04	ug/l	94
16) Acetone	3.83	43	16916	7.09	ug/l	97
21) trans-1,2-Dichloroethene	5.05	96	8563	1.15	ug/l	96
27) cis-1,2-Dichloroethene	6.83	96	255963	30.91	ug/l	92
44) Trichloroethene	8.84	130	320094	36.83	ug/l	100
64) Tetrachloroethene	10.63	164	624177	79.39	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081518\
Data File : VN050652.D
Acq On : 15 Aug 2018 13:59
Operator : MD\SY
Sample : J4465-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
941-MW-02(23.8)

Quant Time: Aug 16 03:19:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

