

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081518\
 Data File : VN050653.D
 Acq On : 15 Aug 2018 14:24
 Operator : MD\SY
 Sample : J4465-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 943-MW-04(23)

Quant Time: Aug 16 13:43:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	614072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	948885	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	851637	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	311081	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	407522	52.65	ug/l	0.00
Spiked Amount						
			Recovery	=	105.30%	
35) Dibromofluoromethane	7.59	113	382511	50.49	ug/l	0.00
Spiked Amount						
			Recovery	=	100.98%	
50) Toluene-d8	10.09	98	1362460	47.79	ug/l	0.00
Spiked Amount						
			Recovery	=	95.58%	
62) 4-Bromofluorobenzene	12.40	95	377861	40.12	ug/l	0.00
Spiked Amount						
			Recovery	=	80.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
64) Tetrachloroethene	10.63	164	3893	0.49	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081518\
Data File : VN050653.D
Acq On : 15 Aug 2018 14:24
Operator : MD\SY
Sample : J4465-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
943-MW-04(23)

Quant Time: Aug 16 13:43:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

