

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081518\
 Data File : VN050663.D
 Acq On : 15 Aug 2018 18:32
 Operator : MD\SY
 Sample : PB1112055ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB1112055ZHE#03

Manual Integrations
 APPROVED

MMDadoda
 8/16/2018 1:18:39 PM

Quant Time: Aug 16 04:13:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	574876	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	968715	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	835996	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	306537	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	403578	55.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.38%	
35) Dibromofluoromethane	7.59	113	371478	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
50) Toluene-d8	10.09	98	1311589	45.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.12%	
62) 4-Bromofluorobenzene	12.40	95	371096	38.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.18%	
Target Compounds						
16) Acetone	3.83	43	12669	5.58	ug/l	92
18) Methyl Acetate	4.33	43	14146m	1.82	ug/l	
43) Isopropyl Acetate	8.17	43	271644	24.45	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081518\
Data File : VN050663.D
Acq On : 15 Aug 2018 18:32
Operator : MD\SY
Sample : PB1112055ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB1112055ZHE#03

Manual Integrations
APPROVED

MMDadoda
8/16/2018 1:18:39 PM

Quant Time: Aug 16 04:13:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

