

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081518\
 Data File : VN050668.D
 Acq On : 15 Aug 2018 21:25
 Operator : MD\SY
 Sample : VN0815WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0815WBL02

Quant Time: Aug 16 04:35:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N072518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 26 15:21:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	113313	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	574482	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	467216	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	233836	30.36	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.20%
60) 4-Bromofluorobenzene	12.40	95	151897	21.16	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	70.53%
63) Toluene-d8	10.09	98	657136	29.45	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.17%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081518\
Data File : VN050668.D
Acq On : 15 Aug 2018 21:25
Operator : MD\SY
Sample : VN0815WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0815WBL02

Quant Time: Aug 16 04:35:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N072518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 26 15:21:50 2018
Response via : Initial Calibration

