

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081518\
 Data File : VN050678.D
 Acq On : 16 Aug 2018 1:57
 Operator : MD\SY
 Sample : PB1112055ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB1112055ZHE#06

Quant Time: Aug 16 05:14:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	612891	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1009141	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	873453	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	311373	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	420679	54.45	ug/l	0.00
Spiked Amount			Recovery	=	108.90%	
35) Dibromofluoromethane	7.59	113	386774	48.01	ug/l	0.00
Spiked Amount			Recovery	=	96.02%	
50) Toluene-d8	10.09	98	1393634	45.96	ug/l	0.00
Spiked Amount			Recovery	=	91.92%	
62) 4-Bromofluorobenzene	12.40	95	387169	38.65	ug/l	0.00
Spiked Amount			Recovery	=	77.30%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	13066	5.35	ug/l	96
18) Methyl Acetate	4.34	43	11612	1.17	ug/l	92
43) Isopropyl Acetate	8.17	43	221965	19.18	ug/l	99
70) Styrene	11.97	104	1576	1.01	ug/l #	80
95) Naphthalene	15.14	128	8750	4.54	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081518\
Data File : VN050678.D
Acq On : 16 Aug 2018 1:57
Operator : MD\SY
Sample : PB1112055ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB1112055ZHE#06

Quant Time: Aug 16 05:14:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

