

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081518\
 Data File : VN050683.D
 Acq On : 16 Aug 2018 4:01
 Operator : MD\SY
 Sample : J4471-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Aug 16 05:24:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	598100	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1002637	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	863453	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	309007	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	415700	55.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.28%	
35) Dibromofluoromethane	7.59	113	388881	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
50) Toluene-d8	10.09	98	1378318	45.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.50%	
62) 4-Bromofluorobenzene	12.40	95	381130	38.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	15216	6.67	ug/l	99
18) Methyl Acetate	4.33	43	13461	1.58	ug/l	98
20) Methylene Chloride	4.55	84	27244	2.60	ug/l	96
43) Isopropyl Acetate	8.17	43	218035	18.96	ug/l	98
95) Naphthalene	15.13	128	76660	9.81	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081518\
Data File : VN050683.D
Acq On : 16 Aug 2018 4:01
Operator : MD\SY
Sample : J4471-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Aug 16 05:24:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

