

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081518\
 Data File : VN050685.D
 Acq On : 16 Aug 2018 4:50
 Operator : MD\SY
 Sample : J4471-24
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-6

Quant Time: Aug 16 08:08:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	610016	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1009093	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	864330	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	309247	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	417364	54.28	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	108.56%	
35) Dibromofluoromethane	7.59	113	388554	48.23	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.46%	
50) Toluene-d8	10.09	98	1382302	45.59	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.18%	
62) 4-Bromofluorobenzene	12.40	95	382485	38.19	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	76.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	14710	6.25	ug/l	98
18) Methyl Acetate	4.33	43	15130	1.84	ug/l	99
20) Methylene Chloride	4.55	84	194241	25.97	ug/l	98
25) 2-Butanone	6.85	43	6665	2.11	ug/l	99
43) Isopropyl Acetate	8.17	43	233419	20.17	ug/l	96
95) Naphthalene	15.14	128	24619	5.77	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081518\
Data File : VN050685.D
Acq On : 16 Aug 2018 4:50
Operator : MD\SY
Sample : J4471-24
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-6

Quant Time: Aug 16 08:08:35 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

