

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081519\
 Data File : VN057299.D
 Acq On : 14 Aug 2019 20:42
 Operator : JC/SP
 Sample : K4231-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-3-D-(2)

Quant Time: Aug 15 08:41:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 08:22:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1156149	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1631184	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1364908	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	368375	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	545045	43.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.74%	
35) Dibromofluoromethane	7.58	113	495284	46.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.64%	
50) Toluene-d8	10.09	98	1890267	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
62) 4-Bromofluorobenzene	12.40	95	525063	37.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.98%	
Target Compounds						
16) Acetone	3.81	43	28629	6.566	ug/l	91
20) Methylene Chloride	4.54	84	73807	6.050	ug/l	99
43) Isopropyl Acetate	8.16	43	160594	5.891	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081519\
Data File : VN057299.D
Acq On : 14 Aug 2019 20:42
Operator : JC/SP
Sample : K4231-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-3-D-(2)

Quant Time: Aug 15 08:41:55 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 08:22:24 2019
Response via : Initial Calibration

