

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081519\
 Data File : VN057303.D
 Acq On : 14 Aug 2019 22:19
 Operator : JC/SP
 Sample : K4324-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T8-E1-(0)

Quant Time: Aug 15 08:48:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 08:22:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1143213	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1607247	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1357442	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	367988	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	537763	43.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.54%	
35) Dibromofluoromethane	7.58	113	481897	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
50) Toluene-d8	10.09	98	1884470	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
62) 4-Bromofluorobenzene	12.40	95	521970	38.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.66%	
Target Compounds						
16) Acetone	3.81	43	48253	11.192	ug/l	96
20) Methylene Chloride	4.55	84	72644	6.022	ug/l	92
43) Isopropyl Acetate	8.16	43	199450	7.886	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081519\
Data File : VN057303.D
Acq On : 14 Aug 2019 22:19
Operator : JC/SP
Sample : K4324-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T8-E1-(0)

Quant Time: Aug 15 08:48:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 08:22:24 2019
Response via : Initial Calibration

