

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081519\
 Data File : VN057310.D
 Acq On : 15 Aug 2019 1:09
 Operator : JC/SP
 Sample : K4325-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20190612-LOC-14

Quant Time: Aug 15 08:59:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 08:22:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1136116	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1598463	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1343968	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	357489	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	540441	43.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.52%	
35) Dibromofluoromethane	7.58	113	475219	45.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.70%	
50) Toluene-d8	10.09	98	1857770	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
62) 4-Bromofluorobenzene	12.40	95	503852	37.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.40%	
Target Compounds						
16) Acetone	3.81	43	26431	6.169	ug/l #	87
18) Methyl Acetate	4.32	43	33082	0.494	ug/l	93
20) Methylene Chloride	4.55	84	367973	30.696	ug/l	98
43) Isopropyl Acetate	8.16	43	295482	12.612	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081519\
Data File : VN057310.D
Acq On : 15 Aug 2019 1:09
Operator : JC/SP
Sample : K4325-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20190612-LOC-14

Quant Time: Aug 15 08:59:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 08:22:24 2019
Response via : Initial Calibration

