

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081519\
 Data File : VN057311.D
 Acq On : 15 Aug 2019 1:33
 Operator : JC/SP
 Sample : K4096-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-081219

Quant Time: Aug 15 09:09:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 08:22:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1130196	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1588087	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1343599	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	364238	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	534718	43.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.04%	
35) Dibromofluoromethane	7.58	113	471909	45.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.66%	
50) Toluene-d8	10.09	98	1861942	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
62) 4-Bromofluorobenzene	12.40	95	508368	37.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.56%	
Target Compounds						
16) Acetone	3.81	43	32069	7.524	ug/l	99
18) Methyl Acetate	4.32	43	32857	0.489	ug/l #	71
20) Methylene Chloride	4.54	84	43275	3.629	ug/l	97
43) Isopropyl Acetate	8.16	43	323294	14.069	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081519\
Data File : VN057311.D
Acq On : 15 Aug 2019 1:33
Operator : JC/SP
Sample : K4096-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-081219

Quant Time: Aug 15 09:09:16 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 08:22:24 2019
Response via : Initial Calibration

