

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081519\
 Data File : VN057312.D
 Acq On : 15 Aug 2019 1:58
 Operator : JC/SP
 Sample : K4088-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-01-081219

Quant Time: Aug 15 09:10:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 08:22:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1123580	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1604134	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1334483	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	360942	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	542461	44.41	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	88.82%	
35) Dibromofluoromethane	7.58	113	479517	45.60	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.20%	
50) Toluene-d8	10.09	98	1863555	48.45	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.90%	
62) 4-Bromofluorobenzene	12.40	95	508361	37.40	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	74.80%	
Target Compounds						
16) Acetone	3.81	43	67862	16.016	ug/l	98
18) Methyl Acetate	4.32	43	33141	0.533	ug/l	97
20) Methylene Chloride	4.55	84	120833	10.192	ug/l	98
43) Isopropyl Acetate	8.16	43	325715	14.028	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081519\
 Data File : VN057312.D
 Acq On : 15 Aug 2019 1:58
 Operator : JC/SP
 Sample : K4088-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleID :
 CL-01-081219

Quant Time: Aug 15 09:10:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 08:22:24 2019
 Response via : Initial Calibration

