

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081519\
 Data File : VN057313.D
 Acq On : 15 Aug 2019 2:22
 Operator : JC/SP
 Sample : K4231-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-3-B-(2)

Quant Time: Aug 15 09:12:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 08:22:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1103441	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1553332	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	1286734	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	340740	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	522425	43.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.10%	
35) Dibromofluoromethane	7.58	113	467693	45.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.86%	
50) Toluene-d8	10.08	98	1803492	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
62) 4-Bromofluorobenzene	12.40	95	484175	36.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.58%	
Target Compounds						
16) Acetone	3.80	43	58757	14.120	ug/l	93
18) Methyl Acetate	4.32	43	34723	0.739	ug/l	96
20) Methylene Chloride	4.54	84	64010	5.498	ug/l	98
43) Isopropyl Acetate	8.16	43	191600	7.828	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081519\
Data File : VN057313.D
Acq On : 15 Aug 2019 2:22
Operator : JC/SP
Sample : K4231-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-3-B-(2)

Quant Time: Aug 15 09:12:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 08:22:24 2019
Response via : Initial Calibration

