

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081519\
 Data File : VN057314.D
 Acq On : 15 Aug 2019 2:46
 Operator : JC/SP
 Sample : K4231-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-3-B-(3)

Quant Time: Aug 15 09:13:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 08:22:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1108065	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1555425	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	1311431	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	330782	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	531414	44.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.24%	
35) Dibromofluoromethane	7.58	113	469991	46.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.20%	
50) Toluene-d8	10.09	98	1807979	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
62) 4-Bromofluorobenzene	12.40	95	479362	36.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.74%	
Target Compounds						
16) Acetone	3.81	43	46240	11.066	ug/l	99
18) Methyl Acetate	4.32	43	31582	0.430	ug/l	98
20) Methylene Chloride	4.55	84	64561	5.522	ug/l	95
43) Isopropyl Acetate	8.16	43	171083	6.789	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081519\
Data File : VN057314.D
Acq On : 15 Aug 2019 2:46
Operator : JC/SP
Sample : K4231-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-3-B-(3)

Quant Time: Aug 15 09:13:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 08:22:24 2019
Response via : Initial Calibration

