


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN081519\
 Data File : VN057288.D
 Acq On : 14 Aug 2019 15:21
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Title : SW846 8260
 Last Update : Thu Aug 15 11:33:14 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.5	14934	PASS
75	95	30	60	48.9	47144	PASS
95	95	100	100	100.0	96373	PASS
96	95	5	9	6.7	6491	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	92.0	88672	PASS
175	174	5	9	8.0	7073	PASS
176	174	95	101	98.2	87032	PASS
177	176	5	9	6.8	5891	PASS