

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081522\
 Data File : VN073906.D
 Acq On : 15 Aug 2022 19:34
 Operator : JC\MD
 Sample : N4152-01
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-02-081022

Quant Time: Aug 15 23:59:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 15 23:50:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.004	168	610664	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.898	114	879076	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	877884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	523608	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	278694	46.856	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.720%
35) Dibromofluoromethane	7.945	113	265410	51.053	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.100%
50) Toluene-d8	10.374	98	933708	46.374	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.740%
62) 4-Bromofluorobenzene	12.674	95	429225	64.234	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	128.460%#
Target Compounds						
18) Methyl Acetate	4.792	43	63781	6.902	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081522\
Data File : VN073906.D
Acq On : 15 Aug 2022 19:34
Operator : JC\MD
Sample : N4152-01
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-02-081022

Quant Time: Aug 15 23:59:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 15 23:50:46 2022
Response via : Initial Calibration

