

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081522\
 Data File : VN073908.D
 Acq On : 15 Aug 2022 20:24
 Operator : JC\MD
 Sample : N4153-01
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-081022

Quant Time: Aug 16 00:00:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 15 23:50:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.004	168	604472	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.898	114	853715	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	831921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	478168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	273383	46.434	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.860%
35) Dibromofluoromethane	7.945	113	257425	50.988	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.980%
50) Toluene-d8	10.380	98	902284	46.144	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.280%
62) 4-Bromofluorobenzene	12.674	95	400851	61.770	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	123.540%#
Target Compounds						
18) Methyl Acetate	4.787	43	68371	7.696	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081522\
Data File : VN073908.D
Acq On : 15 Aug 2022 20:24
Operator : JC\MD
Sample : N4153-01
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-081022

Quant Time: Aug 16 00:00:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 15 23:50:46 2022
Response via : Initial Calibration

