

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081522\
 Data File : VN073916.D
 Acq On : 16 Aug 2022 00:29
 Operator : JC\MD
 Sample : PB146976TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146976TB

Quant Time: Aug 16 04:51:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	656811	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	938072	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	916977	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	495051	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	302761	47.326	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.660%		
35) Dibromofluoromethane	7.951	113	294867	53.152	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	106.300%		
50) Toluene-d8	10.381	98	992644	46.200	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	92.400%		
62) 4-Bromofluorobenzene	12.675	95	426835	59.859	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	119.720%		
Target Compounds						
					Qvalue	
16) Acetone	4.222	43	87819	29.097	ug/l	94
20) Methylene Chloride	5.022	84	25127	3.945	ug/l #	94
37) Ethyl Acetate	7.340	43	79198	8.484	ug/l	97
43) Isopropyl Acetate	8.498	43	326089	24.520	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081522\
Data File : VN073916.D
Acq On : 16 Aug 2022 00:29
Operator : JC\MD
Sample : PB146976TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146976TB

Quant Time: Aug 16 04:51:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

