

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081522\
 Data File : VN073919.D
 Acq On : 16 Aug 2022 01:42
 Operator : JC\MD
 Sample : N4188-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S3

Quant Time: Aug 16 05:13:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	638450	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	917754	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	903035	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	484019	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	299194	48.114	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.220%
35) Dibromofluoromethane	7.951	113	291595	53.725	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.460%
50) Toluene-d8	10.380	98	983513	46.789	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.580%
62) 4-Bromofluorobenzene	12.674	95	415215	59.518	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	119.040%
Target Compounds						
					Qvalue	
16) Acetone	4.222	43	90075	30.703	ug/l	95
20) Methylene Chloride	5.022	84	30805	4.976	ug/l	86
37) Ethyl Acetate	7.339	43	70369	7.705	ug/l	99
43) Isopropyl Acetate	8.498	43	328208	25.226	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081522\
Data File : VN073919.D
Acq On : 16 Aug 2022 01:42
Operator : JC\MD
Sample : N4188-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S3

Quant Time: Aug 16 05:13:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

