

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081522\
 Data File : VN073925.D
 Acq On : 16 Aug 2022 04:09
 Operator : JC\MD
 Sample : N4188-23
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S11

Quant Time: Aug 16 05:16:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	629583	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	921155	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	906586	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	503708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	296738	48.391	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.780%
35) Dibromofluoromethane	7.951	113	292805	53.749	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.500%
50) Toluene-d8	10.380	98	981625	46.527	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.060%
62) 4-Bromofluorobenzene	12.674	95	423557	60.490	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	120.980%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	89386	30.897	ug/l	99
20) Methylene Chloride	5.016	84	32745	5.364	ug/l	97
37) Ethyl Acetate	7.339	43	70473	7.688	ug/l	99
43) Isopropyl Acetate	8.492	43	305323	23.381	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081522\
Data File : VN073925.D
Acq On : 16 Aug 2022 04:09
Operator : JC\MD
Sample : N4188-23
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S11

Quant Time: Aug 16 05:16:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

