

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081522\
 Data File : VN073933.D
 Acq On : 16 Aug 2022 07:24
 Operator : JC\MD
 Sample : N4211-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VOC-WC-1

Quant Time: Aug 16 08:09:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	616623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	896417	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	872816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	440191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	289387	48.184	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.360%
35) Dibromofluoromethane	7.951	113	288621	54.443	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.880%
50) Toluene-d8	10.380	98	956741	46.599	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.200%
62) 4-Bromofluorobenzene	12.669	95	387266	56.834	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.660%
Target Compounds						
					Qvalue	
16) Acetone	4.222	43	57838	20.412	ug/l	94
20) Methylene Chloride	5.016	84	97582	16.320	ug/l	97
37) Ethyl Acetate	7.339	43	72354	8.111	ug/l	99
43) Isopropyl Acetate	8.492	43	312672	24.604	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081522\
Data File : VN073933.D
Acq On : 16 Aug 2022 07:24
Operator : JC\MD
Sample : N4211-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VOC-WC-1

Quant Time: Aug 16 08:09:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

