

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081523\
 Data File : VN078889.D
 Acq On : 15 Aug 2023 15:08
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN081523

Quant Time: Aug 16 02:55:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.694	0.632	8.9	95	-0.01
3 P	Chloromethane	0.969	0.816	15.8	89	0.00
4 C	Vinyl Chloride	0.849	0.800	5.8#	97	0.00
5 T	Bromomethane	0.488	0.454	7.0	101	-0.01
6 T	Chloroethane	0.594	0.516	13.1	93	0.00
7 T	Trichlorofluoromethane	1.035	0.962	7.1	95	-0.01
8 T	Diethyl Ether	0.392	0.364	7.1	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.551	0.523	5.1	97	0.00
10 T	Methyl Iodide	0.617	0.599	2.9	93	-0.01
11 T	Tert butyl alcohol	0.131	0.123	6.1	97	-0.02
12 CM	1,1-Dichloroethene	0.537	0.500	6.9#	95	0.00
13 T	Acrolein	0.067	0.082	-22.4	117	-0.02
14 T	Allyl chloride	1.068	1.044	2.2	95	0.00
15 T	Acrylonitrile	0.354	0.340	4.0	97	0.00
16 T	Acetone	0.321	0.346	-7.8	106	-0.01
17 T	Carbon Disulfide	1.503	1.384	7.9	95	-0.01
18 T	Methyl Acetate	1.356	1.237	8.8	94	-0.01
19 T	Methyl tert-butyl Ether	1.934	1.860	3.8	96	-0.02
20 T	Methylene Chloride	0.684	0.613	10.4	94	-0.01
21 T	trans-1,2-Dichloroethene	0.602	0.553	8.1	93	0.00
22 T	Diisopropyl ether	2.260	2.182	3.5	96	0.00
23 T	Vinyl Acetate	1.488	1.431	3.8	93	0.00
24 P	1,1-Dichloroethane	1.235	1.148	7.0	95	0.00
25 T	2-Butanone	0.515	0.509	1.2	99	0.00
26 T	2,2-Dichloropropane	0.998	1.000	-0.2	100	0.00
27 T	cis-1,2-Dichloroethene	0.732	0.671	8.3	95	0.00
28 T	Bromochloromethane	0.655	0.723	-10.4	110	-0.01
29 T	Tetrahydrofuran	0.341	0.326	4.4	97	0.00
30 C	Chloroform	1.234	1.172	5.0#	95	0.00
31 T	Cyclohexane	1.068	0.942	11.8	96	0.00
32 T	1,1,1-Trichloroethane	1.052	0.993	5.6	95	0.00
33 S	1,2-Dichloroethane-d4	0.778	0.687	11.7	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.337	0.298	11.6	87	0.00
36 T	1,1-Dichloropropene	0.506	0.497	1.8	95	0.00
37 T	Ethyl Acetate	0.566	0.558	1.4	95	0.00
38 T	Carbon Tetrachloride	0.507	0.490	3.4	95	0.00
39 T	Methylcyclohexane	0.466	0.439	5.8	91	0.00
40 TM	Benzene	1.496	1.464	2.1	96	0.00
41 T	Methacrylonitrile	0.309	0.314	-1.6	97	-0.01
42 TM	1,2-Dichloroethane	0.559	0.536	4.1	96	0.00
43 T	Isopropyl Acetate	1.001	0.936	6.5	93	0.00
44 TM	Trichloroethene	0.356	0.340	4.5	95	0.00
45 C	1,2-Dichloropropane	0.398	0.393	1.3#	96	0.00
46 T	Dibromomethane	0.273	0.260	4.8	97	0.00
47 T	Bromodichloromethane	0.540	0.539	0.2	99	0.00
48 T	Methyl methacrylate	0.454	0.442	2.6	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081523\
 Data File : VN078889.D
 Acq On : 15 Aug 2023 15:08
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN081523

Quant Time: Aug 16 02:55:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	98	0.00
50 S	Toluene-d8	1.250	1.170	6.4	90	0.00
51 T	4-Methyl-2-Pentanone	0.594	0.582	2.0	95	0.00
52 CM	Toluene	0.919	0.912	0.8#	97	0.00
53 T	t-1,3-Dichloropropene	0.579	0.586	-1.2	96	0.00
54 T	cis-1,3-Dichloropropene	0.630	0.622	1.3	96	0.00
55 T	1,1,2-Trichloroethane	0.368	0.358	2.7	96	0.00
56 T	Ethyl methacrylate	0.571	0.592	-3.7	98	0.00
57 T	1,3-Dichloropropane	0.659	0.646	2.0	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.274	0.251	8.4	89	0.00
59 T	2-Hexanone	0.430	0.440	-2.3	98	0.00
60 T	Dibromochloromethane	0.388	0.387	0.3	98	0.00
61 T	1,2-Dibromoethane	0.374	0.374	0.0	99	0.00
62 S	4-Bromofluorobenzene	0.408	0.401	1.7	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.314	0.291	7.3	98	0.00
65 PM	Chlorobenzene	1.064	1.022	3.9	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.387	3.7	96	0.00
67 C	Ethyl Benzene	1.935	1.871	3.3#	98	0.00
68 T	m/p-Xylenes	0.724	0.710	1.9	97	0.00
69 T	o-Xylene	0.704	0.689	2.1	95	0.00
70 T	Styrene	1.174	1.183	-0.8	97	0.00
71 P	Bromoform	0.292	0.290	0.7	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	4.169	3.823	8.3	97	0.00
74 T	N-amyl acetate	2.104	1.906	9.4	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.505	1.304	13.4	96	0.00
76 T	1,2,3-Trichloropropane	1.205	1.065	11.6	97	0.00
77 T	Bromobenzene	1.002	0.885	11.7	94	0.00
78 T	n-propylbenzene	4.834	4.546	6.0	97	0.00
79 T	2-Chlorotoluene	2.999	2.713	9.5	98	0.00
80 T	1,3,5-Trimethylbenzene	3.304	3.086	6.6	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.472	0.436	7.6	95	0.00
82 T	4-Chlorotoluene	2.934	2.691	8.3	95	0.00
83 T	tert-Butylbenzene	2.776	2.487	10.4	92	0.00
84 T	1,2,4-Trimethylbenzene	3.271	3.108	5.0	98	0.00
85 T	sec-Butylbenzene	3.820	3.511	8.1	96	0.00
86 T	p-Isopropyltoluene	3.052	2.887	5.4	96	0.00
87 T	1,3-Dichlorobenzene	1.724	1.581	8.3	94	0.00
88 T	1,4-Dichlorobenzene	1.726	1.616	6.4	98	0.00
89 T	n-Butylbenzene	2.512	2.393	4.7	95	0.00
90 T	Hexachloroethane	0.612	0.535	12.6	97	0.00
91 T	1,2-Dichlorobenzene	1.715	1.586	7.5	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.284	0.235	17.3	93	0.00
93 T	1,2,4-Trichlorobenzene	0.608	0.576	5.3	93	0.00
94 T	Hexachlorobutadiene	0.329	0.268	18.5	90	0.00
95 T	Naphthalene	2.054	1.975	3.8	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081523\
Data File : VN078889.D
Acq On : 15 Aug 2023 15:08
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN081523

Quant Time: Aug 16 02:55:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.666	0.622	6.6	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6