

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081524\
 Data File : VN083326.D
 Acq On : 15 Aug 2024 13:10
 Operator : JC\MD
 Sample : P3617-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Aug 16 01:50:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	128329	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	251863	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	258981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	114574	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	106675	58.400	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.800%
35) Dibromofluoromethane	8.171	113	83676	53.226	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.460%
50) Toluene-d8	10.571	98	313083	53.389	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.780%
62) 4-Bromofluorobenzene	12.853	95	128174	56.064	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	112.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081524\
Data File : VN083326.D
Acq On : 15 Aug 2024 13:10
Operator : JC\MD
Sample : P3617-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Aug 16 01:50:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

