

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081524\
 Data File : VN083331.D
 Acq On : 15 Aug 2024 15:11
 Operator : JC\MD
 Sample : P3620-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 920-J-WPO-0.25-081424

Quant Time: Aug 16 01:52:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	124063	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	243725	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	255138	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	114799	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	103661	58.702	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.400%
35) Dibromofluoromethane	8.171	113	82034	53.924	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.840%
50) Toluene-d8	10.565	98	307673	54.218	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.440%
62) 4-Bromofluorobenzene	12.847	95	126756	57.295	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	114.600%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	7163	10.366	ug/l	95
29) Tetrahydrofuran	7.841	42	23316	33.995	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081524\
Data File : VN083331.D
Acq On : 15 Aug 2024 15:11
Operator : JC\MD
Sample : P3620-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
920-J-WPO-0.25-081424

Quant Time: Aug 16 01:52:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

