

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081524\
 Data File : VN083337.D
 Acq On : 15 Aug 2024 17:37
 Operator : JC\MD
 Sample : P3456-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-WATER-REF-4

Quant Time: Aug 16 03:58:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	117857	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	236930	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	252402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116043	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	102649	61.189	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 122.380%	
35) Dibromofluoromethane	8.165	113	81581	55.165	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.320%	
50) Toluene-d8	10.565	98	303750	55.062	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.120%	
62) 4-Bromofluorobenzene	12.847	95	127241	59.164	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 118.320%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081524\
Data File : VN083337.D
Acq On : 15 Aug 2024 17:37
Operator : JC\MD
Sample : P3456-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-WATER-REF-4

Quant Time: Aug 16 03:58:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

