

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081525\
 Data File : VN087574.D
 Acq On : 15 Aug 2025 10:51
 Operator : JC\MD
 Sample : VN0815MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0815MBL01

Quant Time: Aug 18 05:40:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	228978	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.082	114	520207	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	474017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	214374	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	213880	55.049	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.100%
35) Dibromofluoromethane	8.153	113	161693	45.060	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.120%
50) Toluene-d8	10.547	98	580394	45.343	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.680%
62) 4-Bromofluorobenzene	12.829	95	197517	41.767	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	83.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081525\
Data File : VN087574.D
Acq On : 15 Aug 2025 10:51
Operator : JC\MD
Sample : VN0815MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0815MBL01

Quant Time: Aug 18 05:40:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

