

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081525\
 Data File : VN087575.D
 Acq On : 15 Aug 2025 11:12
 Operator : JC\MD
 Sample : VN0815WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0815WBL01

Quant Time: Aug 18 05:40:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	192276	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	377502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	347368	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	154069	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	196309	60.171	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	120.340%
35) Dibromofluoromethane	8.147	113	131619	50.545	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.080%
50) Toluene-d8	10.547	98	471339	50.743	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.480%
62) 4-Bromofluorobenzene	12.829	95	164464	47.924	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081525\
Data File : VN087575.D
Acq On : 15 Aug 2025 11:12
Operator : JC\MD
Sample : VN0815WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0815WBL01

Quant Time: Aug 18 05:40:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

