

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081525\
 Data File : VN087582.D
 Acq On : 15 Aug 2025 15:19
 Operator : JC\MD
 Sample : Q2864-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LAW-25-122-126

Quant Time: Aug 18 05:44:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	283272	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	641823	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	589000	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	285041	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	272912	56.780	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.560%
35) Dibromofluoromethane	8.153	113	196879	44.469	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.940%
50) Toluene-d8	10.547	98	725706	45.952	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.900%
62) 4-Bromofluorobenzene	12.829	95	277581	47.575	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.140%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	475520	211.001	ug/l	96
52) Toluene	10.612	92	5950	0.518	ug/l	78
73) Isopropylbenzene	12.676	105	7584	0.423	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081525\
Data File : VN087582.D
Acq On : 15 Aug 2025 15:19
Operator : JC\MD
Sample : Q2864-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LAW-25-122-126

Quant Time: Aug 18 05:44:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

