

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081525\
 Data File : VN087588.D
 Acq On : 15 Aug 2025 17:29
 Operator : JC\MD
 Sample : Q2844-11
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MOO-25-0228

Quant Time: Aug 18 05:48:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	247646	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	548462	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	463252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.776	152	209456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	220599	52.498	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	105.000%	
35) Dibromofluoromethane	8.147	113	163312	43.167	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	86.340%	
50) Toluene-d8	10.547	98	601560	44.575	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	89.160%	
62) 4-Bromofluorobenzene	12.829	95	190375	38.182	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	76.360%#	
Target Compounds						
52) Toluene	10.612	92	45018	4.585	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081525\
Data File : VN087588.D
Acq On : 15 Aug 2025 17:29
Operator : JC\MD
Sample : Q2844-11
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MOO-25-0228

Quant Time: Aug 18 05:48:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

