

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081618\
 Data File : VN050709.D
 Acq On : 16 Aug 2018 18:25
 Operator : MD\SY
 Sample : J4269-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Aug 17 04:33:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	578511	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	979563	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	837603	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	303515	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	405011	55.54	ug/l	0.00
Spiked Amount			Recovery	=	111.08%	
35) Dibromofluoromethane	7.59	113	376117	48.09	ug/l	0.00
Spiked Amount			Recovery	=	96.18%	
50) Toluene-d8	10.09	98	1331147	45.23	ug/l	0.00
Spiked Amount			Recovery	=	90.46%	
62) 4-Bromofluorobenzene	12.40	95	368425	37.89	ug/l	0.00
Spiked Amount			Recovery	=	75.78%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.06	50	2216	Below Cal		98
5) Bromomethane	2.57	94	2628	Below Cal		82
6) Chloroethane	2.71	64	203	Below Cal	#	43
10) Methyl Iodide	3.96	142	533	7.043	ug/l #	42
11) Tert butyl alcohol	4.80	59	1774	4.455	ug/l #	72
14) Allyl chloride	4.56	41	18949	1.941	ug/l #	25
16) Acetone	3.83	43	40791	21.076	ug/l	98
18) Methyl Acetate	4.33	43	18650	2.700	ug/l #	79
20) Methylene Chloride	4.55	84	509430	74.112	ug/l	98
25) 2-Butanone	6.86	43	5272	1.763	ug/l	93
29) Tetrahydrofuran	7.23	42	792	0.507	ug/l #	39
30) Chloroform	7.38	83	7118	0.542	ug/l	98
31) Cyclohexane	7.67	56	10188	0.580	ug/l #	1
36) 1,1-Dichloropropene	7.67	75	39783	3.648	ug/l #	48
38) Carbon Tetrachloride	7.67	117	52850	4.672	ug/l #	17
43) Isopropyl Acetate	8.17	43	286180	25.477	ug/l #	91
48) Methyl methacrylate	9.19	41	7562	1.338	ug/l #	20
51) 4-Methyl-2-Pentanone	9.99	43	2315	0.305	ug/l	93
56) Ethyl methacrylate	10.35	69	63	1.911	ug/l #	1
59) 2-Hexanone	10.77	43	16200	3.304	ug/l #	57
70) Styrene	11.97	104	73	0.944	ug/l #	58
71) Bromoform	12.40	173	3464	0.669	ug/l #	1
75) 1,1,2,2-Tetrachloroethane	12.41	83	97	Below Cal	#	1
76) 1,2,3-Trichloropropane	12.40	75	174967	35.380	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	174967	128.161	ug/l #	13
95) Naphthalene	15.14	128	45559	7.462	ug/l	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081618\
Data File : VN050709.D
Acq On : 16 Aug 2018 18:25
Operator : MD\SY
Sample : J4269-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Aug 17 04:33:05 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1891

Delta R.T. 0.00 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:168 Resp: 578511

Ion Ratio Lower Upper




168 100

99 49.3 40.8 61.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#3

Chloromethane

Concen: Below Cal

RT: 2.06 min Scan# 145

Delta R.T. 0.00 min

Lab File: VN050709.D


Acq: 16 Aug 2018 18:25

Tgt Ion: 50 Resp: 2216

Ion Ratio Lower Upper



50 100


52 31.2 26.0 39.0



Abundance Ion 49.90 (49.60 to 50.60): VNO

Ion 51.90 (51.60 to 52.60): VNO





#5

Bromomethane

Concen: Below Cal

RT: 2.57 min Scan# 304

Delta R.T. 0.01 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 94 Resp: 2628

Ion Ratio Lower Upper




94 100

96 75.4 74.0 111.0



Abundance Ion 93.90 (93.60 to 94.60): VNO

Ion 95.90 (95.60 to 96.60): VNO



#6

Chloroethane

Concen: Below Cal

RT: 2.71 min Scan# 349

Delta R.T. 0.01 min

Lab File: VN050709.D


Acq: 16 Aug 2018 18:25

Tgt Ion: 64 Resp: 203

Ion Ratio Lower Upper



64 100


66 0.0 25.7 38.5#



Abundance Ion 63.90 (63.60 to 64.60): VNO

Ion 65.90 (65.60 to 66.60): VNO









#10
Methyl Iodide
Concen: 7.043 ug/l
RT: 3.96 min Scan# 736
Delta R.T. 0.00 min
Lab File: VN050709.D
Acq: 16 Aug 2018 18:25




Instrument :
MSVOA_N
ClientSampled :


Tot Ion:142 Resp: 533
Ion Ratio Lower Upper
142 100
127 0.0 32.6 49.0#
141 0.0 11.5 17.3#



#11
Tert butyl alcohol
Concen: 4.455 ug/l
RT: 4.80 min Scan# 998
Delta R.T. 0.01 min
Lab File: VN050709.D
Acq: 16 Aug 2018 18:25

Tgt Ion: 59 Resp: 1774
Ion Ratio Lower Upper
59 100
57 0.0 8.4 12.6#









#14
Allvl chloride
Concen: 1.941 ug/l
RT: 4.56 min Scan# 923
Delta R.T. 0.23 min
Lab File: VN050709.D
Acq: 16 Aug 2018 18:25




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 41 Resp: 18949
Ion Ratio Lower Upper
41 100
39 0.0 51.4 77.0#
76 0.0 29.4 44.0#



#16
Acetone
Concen: 21.076 ug/l
RT: 3.83 min Scan# 696
Delta R.T. 0.01 min
Lab File: VN050709.D
Acq: 16 Aug 2018 18:25

Tgt Ion: 43 Resp: 40791
Ion Ratio Lower Upper
43 100
58 32.5 27.1 40.7









#18
Methvl Acetate
Concen: 2.700 ug/l
RT: 4.33 min Scan# 853
Delta R.T. 0.01 min
Lab File: VN050709.D
Acq: 16 Aug 2018 18:25




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 43 Resp: 18650
Ion Ratio Lower Upper
43 100
74 14.2 19.7 29.5#



#20
Methylene Chloride
Concen: 74.112 ug/l
RT: 4.55 min Scan# 922
Delta R.T. 0.00 min
Lab File: VN050709.D
Acq: 16 Aug 2018 18:25

Tgt Ion: 84 Resp: 509430
Ion Ratio Lower Upper
84 100
49 119.5 92.6 138.8
51 36.9 28.6 43.0
86 64.8 52.2 78.2





#25

2-Butanone

Concen: 1.763 ug/l

RT: 6.86 min Scan# 1640

Delta R.T. 0.02 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25





Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 43 Resp: 5272

Ion Ratio Lower Upper

43 100

72 23.5 21.8 32.6



#29

Tetrahydrofuran

Concen: 0.507 ug/l

RT: 7.23 min Scan# 1754

Delta R.T. 0.02 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25




Tgt Ion: 42 Resp: 792


Ion Ratio Lower Upper

42 100

72 0.0 35.8 53.6#

71 8.7 33.4 50.0#









#30
Chloroform
Concen: 0.542 ug/l
RT: 7.38 min Scan# 1800
Delta R.T. 0.00 min
Lab File: VN050709.D
Acq: 16 Aug 2018 18:25




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 83 Resp: 7118
Ion Ratio Lower Upper
83 100
85 67.4 52.5 78.7



#31
Cyclohexane
Concen: 0.580 ug/l
RT: 7.67 min Scan# 1891
Delta R.T. 0.02 min
Lab File: VN050709.D
Acq: 16 Aug 2018 18:25

Tgt Ion: 56 Resp: 10188
Ion Ratio Lower Upper
56 100
69 178.5 25.8 38.6#
84 136.5 67.8 101.6#









#33
 1,2-Dichloroethane-d4
 Concen: 55.541 ug/l
 RT: 8.03 min Scan# 2003
 Delta R.T. 0.00 min
 Lab File: VN050709.D
 Acq: 16 Aug 2018 18:25




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 65 Resp: 405011
 Ion Ratio Lower Upper
 65 100
 67 54.0 0.0 109.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2177
 Delta R.T. 0.00 min
 Lab File: VN050709.D
 Acq: 16 Aug 2018 18:25

Tgt Ion: 114 Resp: 979563
 Ion Ratio Lower Upper
 114 100
 63 19.4 0.0 40.0
 88 15.2 0.0 30.8









#35
 Dibromofluoromethane
 Concen: 48.094 ug/l
 RT: 7.59 min Scan# 1867
 Delta R.T. 0.00 min
 Lab File: VN050709.D
 Acq: 16 Aug 2018 18:25




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion:113 Resp: 376117
 Ion Ratio Lower Upper
 113 100
 111 102.5 81.0 121.6
 192 22.5 17.6 26.4



#36
 1,1-Dichloropropene
 Concen: 3.648 ug/l
 RT: 7.67 min Scan# 1890
 Delta R.T. -0.13 min
 Lab File: VN050709.D
 Acq: 16 Aug 2018 18:25

Tgt Ion: 75 Resp: 39783
 Ion Ratio Lower Upper
 75 100
 110 7.6 18.3 54.9#
 77 0.0 25.0 37.4#





#38

Carbon Tetrachloride

Concen: 4.672 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. -0.11 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25

Instrument :

MSVOA_N

ClientSampled :


Tot Ion:117 Resp: 52850

Ion Ratio Lower Upper

117 100

119 6.5 76.6 115.0#




121 0.0 25.0 37.6#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 25.477 ug/l

RT: 8.17 min Scan# 2047

Delta R.T. 0.01 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25


Tgt Ion: 43 Resp: 286180

Ion Ratio Lower Upper

43 100

61 15.7 16.2 24.2#



87 10.2 10.9 16.3#




Abundance Ion 43.00 (42.70 to 43.70): V

Ion 61.00 (60.70 to 61.70): V

Ion 87.00 (86.70 to 87.70): V









#48
Methvl methacrylate
Concen: 1.338 ug/l
RT: 9.19 min Scan# 2363
Delta R.T. -0.01 min
Lab File: VN050709.D
Acq: 16 Aug 2018 18:25

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 41 Resp: 7562
Ion Ratio Lower Upper
41 100
69 0.0 73.4 110.0#
39 16.2 43.0 64.6#




Abundance Ion 41.00 (40.70 to 41.70): VNC
Ion 69.00 (68.70 to 69.70): VNC
Ion 39.00 (38.70 to 39.70): VNC






#50
Toluene-d8
Concen: 45.227 ug/l
RT: 10.09 min Scan# 2644
Delta R.T. 0.00 min
Lab File: VN050709.D
Acq: 16 Aug 2018 18:25

Tgt Ion: 98 Resp: 1331147
Ion Ratio Lower Upper
98 100
100 63.0 51.8 77.8



Abundance Ion 98.00 (97.70 to 98.70): VNC
Ion 100.00 (99.70 to 100.70): VNC





#51

4-Methyl-2-Pentanone

Concen: 0.305 ug/l

RT: 9.99 min Scan# 2613

Delta R.T. 0.01 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 43 Resp: 2315

Ion Ratio Lower Upper




43 100

58 36.5 32.5 48.7



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#56

Ethyl methacrylate

Concen: 1.911 ug/l

RT: 10.35 min Scan# 2725

Delta R.T. -0.08 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25


Tgt Ion: 69 Resp: 63

Ion Ratio Lower Upper

69 100

41 309.5 49.7 74.5#



39 0.0 24.2 36.2#




Abundance Ion 69.00 (68.70 to 69.70): VN0

Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0





#59

2-Hexanone

Concen: 3.304 ug/l

RT: 10.77 min Scan# 2854

Delta R.T. 0.01 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25

Instrument :

MSVOA_N





ClientSampleId :

Tot Ion: 43 Resp: 16200

Ion Ratio Lower Upper

43 100

58 24.6 28.0 84.0#



#62

4-Bromofluorobenzene

Concen: 37.890 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.00 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25




Tgt Ion: 95 Resp: 368425


Ion Ratio Lower Upper

95 100

174 90.6 0.0 177.8

176 86.8 0.0 175.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 3054
Delta R.T. 0.00 min
Lab File: VN050709.D
Acq: 16 Aug 2018 18:25

Instrument :
MSVOA_N
ClientSampled :

Tot Ion:117 Resp: 837603
Ion Ratio Lower Upper
117 100
82 52.6 42.4 63.6
119 31.7 25.8 38.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VN
Ion 118.90 (118.60 to 119.60): V






#70
Styrene
Concen: 0.944 ug/l
RT: 11.97 min Scan# 3228
Delta R.T. 0.01 min
Lab File: VN050709.D
Acq: 16 Aug 2018 18:25

Tgt Ion:104 Resp: 73
Ion Ratio Lower Upper
104 100
78 0.0 39.1 58.7#
103 43.8 44.9 67.3#



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VN
Ion 103.00 (102.70 to 103.70): V





#71

Bromoform

Concen: 0.669 ug/l

RT: 12.40 min Scan# 3362

Delta R.T. 0.27 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:173 Resp: 3464

Ion Ratio Lower Upper

173 100

175 593.4 24.4 73.2#




254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3656

Delta R.T. 0.00 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25


Tgt Ion:152 Resp: 303515

Ion Ratio Lower Upper

152 100

115 56.9 28.1 84.2



150 156.9 0.0 347.8




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#75

1,1,2,2-Tetrachloroethane

Concen: Below Cal

RT: 12.41 min Scan# 3365

Delta R.T. -0.09 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25

Instrument :

MSVOA_N

ClientSampleId :





Tot Ion: 83 Resp: 97

Ion Ratio Lower Upper

83 100

131 420.6 5.3 15.9#

85 0.0 32.1 96.5#



#76

1,2,3-Trichloropropane

Concen: 35.380 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.15 min

Lab File: VN050709.D




Acq: 16 Aug 2018 18:25


Tgt Ion: 75 Resp: 174967

Ion Ratio Lower Upper

75 100

77 1.2 0.0 0.0#





#81

trans-1,4-Dichloro-2-butene

Concen: 128.161 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.10 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25

Instrument :

MSVOA_N

ClientSampled :


Tot Ion: 75 Resp: 174967

Ion Ratio Lower Upper

75 100

53 0.2 72.2 108.2#




89 0.0 36.3 54.5#



Abundance Ion 74.90 (74.60 to 75.60): VN050709.D (-3363) (-)

Ion 53.00 (52.70 to 53.70): VN050709.D (-3363) (-)

Ion 88.90 (88.60 to 89.60): VN050709.D (-3363) (-)



#95

Naphthalene

Concen: 7.462 ug/l

RT: 15.14 min Scan# 4213

Delta R.T. 0.00 min

Lab File: VN050709.D

Acq: 16 Aug 2018 18:25


Tgt Ion:128 Resp: 45559

Ion Ratio Lower Upper

128 100

127 13.0 10.3 15.5

129 10.8 8.5 12.7



Abundance Ion 128.00 (127.70 to 128.70): VN050709.D (-4212) (-)

Ion 127.00 (126.70 to 127.70): VN050709.D (-4212) (-)

Ion 129.00 (128.70 to 129.70): VN050709.D (-4212) (-)

