

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081618\
 Data File : VN050700.D
 Acq On : 16 Aug 2018 14:41
 Operator : MD\SY
 Sample : J4502-03DL 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BROWN-FRAC-TANKDL

Quant Time: Aug 17 04:21:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	629479	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1026926	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	849159	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	305023	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	427519	53.88	ug/l	0.00
Spiked Amount			Recovery	=	107.76%	
35) Dibromofluoromethane	7.59	113	401312	48.95	ug/l	0.00
Spiked Amount			Recovery	=	97.90%	
50) Toluene-d8	10.09	98	1393271	45.15	ug/l	0.00
Spiked Amount			Recovery	=	90.30%	
62) 4-Bromofluorobenzene	12.40	95	372355	36.53	ug/l	0.00
Spiked Amount			Recovery	=	73.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.55	84	483152	64.43	ug/l	98
52) Toluene	10.16	92	32420	1.56	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081618\
Data File : VN050700.D
Acq On : 16 Aug 2018 14:41
Operator : MD\SY
Sample : J4502-03DL 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BROWN-FRAC-TANKDL

Quant Time: Aug 17 04:21:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

