

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081618\
 Data File : VN050701.D
 Acq On : 16 Aug 2018 15:06
 Operator : MD\SY
 Sample : PB112098ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112098ZHE#05

Quant Time: Aug 17 04:25:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	621456	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1053360	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	905034	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	321771	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	430760	54.99	ug/l	0.00
Spiked Amount			Recovery	=	109.98%	
35) Dibromofluoromethane	7.59	113	401323	47.72	ug/l	0.00
Spiked Amount			Recovery	=	95.44%	
50) Toluene-d8	10.09	98	1422652	44.95	ug/l	0.00
Spiked Amount			Recovery	=	89.90%	
62) 4-Bromofluorobenzene	12.40	95	393150	37.60	ug/l	0.00
Spiked Amount			Recovery	=	75.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	21356	9.52	ug/l	98
20) Methylene Chloride	4.55	84	15848	0.88	ug/l	94
43) Isopropyl Acetate	8.17	43	274183	22.70	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081618\
Data File : VN050701.D
Acq On : 16 Aug 2018 15:06
Operator : MD\SY
Sample : PB112098ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112098ZHE#05

Quant Time: Aug 17 04:25:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

