

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081618\
 Data File : VN050711.D
 Acq On : 16 Aug 2018 19:14
 Operator : MD\SY
 Sample : J4269-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AU-06-081518-A

Quant Time: Aug 17 05:13:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	577510	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	982561	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	823286	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	284259	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	407396	55.96	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	111.92%	
35) Dibromofluoromethane	7.59	113	375549	47.87	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.74%	
50) Toluene-d8	10.09	98	1335517	45.24	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	90.48%	
62) 4-Bromofluorobenzene	12.40	95	356298	36.53	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	73.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	23268	11.42	ug/l	95
20) Methylene Chloride	4.56	84	20328	1.71	ug/l	98
30) Chloroform	7.37	83	27274	2.08	ug/l	92
43) Isopropyl Acetate	8.17	43	269054	23.88	ug/l	93
47) Bromodichloromethane	9.40	83	5943	0.53	ug/l	87
95) Naphthalene	15.14	128	32961	6.64	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081618\
Data File : VN050711.D
Acq On : 16 Aug 2018 19:14
Operator : MD\SY
Sample : J4269-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AU-06-081518-A

Quant Time: Aug 17 05:13:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

