

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN081618\
 Data File : VN050689.D
 Acq On : 16 Aug 2018 9:06
 Operator : MD\SY
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 17 02:58:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.564	0.505	10.5	93	0.00
3 P	Chloromethane	0.769	0.649	15.6	97	0.00
4 C	Vinyl Chloride	0.749	0.668	10.8#	94	0.00
5 T	Bromomethane	0.484	0.385	20.5#	94	0.00
6 T	Chloroethane	0.470	0.408	13.2	97	0.00
7 T	Trichlorofluoromethane	0.983	0.883	10.2	97	0.00
8 T	Diethyl Ether	0.327	0.299	8.6	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.594	0.560	5.7	101	0.00
10 T	Methyl Iodide	0.384	0.407	-6.0	111	0.00
11 T	Tert butyl alcohol	0.034	0.031	8.8	93	0.00
12 CM	1,1-Dichloroethene	0.543	0.486	10.5#	94	0.00
13 T	Acrolein	0.014	0.009	35.7#	70	0.00
14 T	Allyl chloride	0.844	0.791	6.3	98	0.00
15 T	Acrylonitrile	0.189	0.176	6.9	95	0.00
16 T	Acetone	0.174	0.161	7.5	104	0.00
17 T	Carbon Disulfide	1.707	1.521	10.9	95	0.00
18 T	Methyl Acetate	0.591	0.422	28.6#	98	0.00
19 T	Methyl tert-butyl Ether	1.381	1.322	4.3	95	0.00
20 T	Methylene Chloride	0.685	0.583	14.9	97	0.00
21 T	trans-1,2-Dichloroethene	0.588	0.547	7.0	97	0.00
22 T	Diisopropyl ether	1.730	1.693	2.1	95	0.00
23 T	Vinyl Acetate	1.132	1.116	1.4	95	0.00
24 P	1,1-Dichloroethane	1.120	1.010	9.8	97	0.00
25 T	2-Butanone	0.258	0.253	1.9	100	0.00
26 T	2,2-Dichloropropane	0.749	0.841	-12.3	123	0.00
27 T	cis-1,2-Dichloroethene	0.655	0.611	6.7	97	0.00
28 T	Bromochloromethane	0.510	0.491	3.7	100	0.00
29 T	Tetrahydrofuran	0.135	0.129	4.4	95	0.00
30 C	Chloroform	1.135	1.017	10.4#	97	0.00
31 T	Cyclohexane	1.095	0.920	16.0	97	0.00
32 T	1,1,1-Trichloroethane	0.955	0.867	9.2	96	0.00
33 S	1,2-Dichloroethane-d4	0.630	0.585	7.1	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.399	0.401	-0.5	94	0.00
36 T	1,1-Dichloropropene	0.557	0.586	-5.2	99	0.00
37 T	Ethyl Acetate	0.315	0.327	-3.8	96	0.00
38 T	Carbon Tetrachloride	0.577	0.572	0.9	98	0.00
39 T	Methylcyclohexane	0.576	0.650	-12.8	100	0.00
40 TM	Benzene	1.693	1.730	-2.2	97	0.00
41 T	Methacrylonitrile	0.170	0.159	6.5	84	0.00
42 TM	1,2-Dichloroethane	0.519	0.521	-0.4	97	0.00
43 T	Isopropyl Acetate	0.573	0.569	0.7	94	0.00
44 TM	Trichloroethene	0.454	0.452	0.4	96	0.00
45 C	1,2-Dichloropropane	0.451	0.457	-1.3#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN081618\
 Data File : VN050689.D
 Acq On : 16 Aug 2018 9:06
 Operator : MD\SY
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 17 02:58:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.266	0.265	0.4	99	0.00
47 T	Bromodichloromethane	0.569	0.566	0.5	96	0.00
48 T	Methyl methacrylate	0.289	0.286	1.0	94	0.00
49 T	1,4-Dioxane	0.004	0.004	0.0	97	0.00
50 S	Toluene-d8	1.502	1.543	-2.7	96	0.00
51 T	4-Methyl-2-Pentanone	0.388	0.409	-5.4	96	0.00
52 CM	Toluene	1.011	1.087	-7.5#	99	0.00
53 T	t-1,3-Dichloropropene	0.530	0.578	-9.1	101	0.00
54 T	cis-1,3-Dichloropropene	0.608	0.666	-9.5	100	0.00
55 T	1,1,2-Trichloroethane	0.378	0.381	-0.8	97	0.00
56 T	Ethyl methacrylate	0.427	0.471	-10.3	96	0.00
57 T	1,3-Dichloropropane	0.619	0.637	-2.9	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.200	0.225	-12.5	95	0.00
59 T	2-Hexanone	0.250	0.280	-12.0	100	0.00
60 T	Dibromochloromethane	0.417	0.428	-2.6	97	0.00
61 T	1,2-Dibromoethane	0.360	0.372	-3.3	97	0.00
62 S	4-Bromofluorobenzene	0.496	0.519	-4.6	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.464	0.445	4.1	94	0.00
65 PM	Chlorobenzene	1.244	1.253	-0.7	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.466	0.459	1.5	97	0.00
67 C	Ethyl Benzene	2.009	2.169	-8.0#	99	0.00
68 T	m/p-Xylenes	0.768	0.848	-10.4	99	0.00
69 T	o-Xylene	0.733	0.794	-8.3	98	0.00
70 T	Styrene	1.177	1.339	-13.8	101	0.00
71 P	Bromoform	0.309	0.313	-1.3	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.907	3.958	-1.3	100	0.00
74 T	N-amyl acetate	1.016	0.987	2.9	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.068	0.925	13.4	99	0.00
76 T	1,2,3-Trichloropropane	0.899	1.122	-24.8#	146	0.00
77 T	Bromobenzene	1.061	1.005	5.3	99	0.00
78 T	n-propylbenzene	4.356	4.612	-5.9	102	0.00
79 T	2-Chlorotoluene	2.716	2.701	0.6	100	0.00
80 T	1,3,5-Trimethylbenzene	3.128	3.304	-5.6	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.225	0.241	-7.1	110	0.00
82 T	4-Chlorotoluene	2.684	2.772	-3.3	101	0.00
83 T	tert-Butylbenzene	2.715	2.783	-2.5	102	0.00
84 T	1,2,4-Trimethylbenzene	3.139	3.369	-7.3	101	0.00
85 T	sec-Butylbenzene	3.548	3.780	-6.5	103	0.00
86 T	p-Isopropyltoluene	3.006	3.333	-10.9	103	0.00
87 T	1,3-Dichlorobenzene	1.806	1.798	0.4	101	0.00
88 T	1,4-Dichlorobenzene	1.792	1.757	2.0	102	0.00
89 T	n-Butylbenzene	2.428	2.743	-13.0	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN081618\
Data File : VN050689.D
Acq On : 16 Aug 2018 9:06
Operator : MD\SY
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 17 02:58:35 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.605	0.580	4.1	106	0.00
91 T	1,2-Dichlorobenzene	1.762	1.714	2.7	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.139	0.132	5.0	99	0.00
93 T	1,2,4-Trichlorobenzene	0.767	0.855	-11.5	100	0.00
94 T	Hexachlorobutadiene	0.540	0.540	0.0	109	0.00
95 T	Naphthalene	1.531	1.645	-7.4	93	0.00
96 T	1,2,3-Trichlorobenzene	0.745	0.812	-9.0	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6