

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081619\
 Data File : VN057329.D
 Acq On : 15 Aug 2019 14:15
 Operator : JC/SP
 Sample : K4264-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-3-C1-(3)

Quant Time: Aug 16 02:15:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1116526	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1554483	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	1262941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	321532	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	515690	42.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.98%	
35) Dibromofluoromethane	7.58	113	466728	45.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.60%	
50) Toluene-d8	10.08	98	1797846	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
62) 4-Bromofluorobenzene	12.40	95	460513	34.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.92%	
Target Compounds						
16) Acetone	3.81	43	23739	5.638	ug/l	97
20) Methylene Chloride	4.54	84	18673	1.585	ug/l	98
43) Isopropyl Acetate	8.16	43	155569	6.588	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081619\
Data File : VN057329.D
Acq On : 15 Aug 2019 14:15
Operator : JC/SP
Sample : K4264-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-3-C1-(3)

Quant Time: Aug 16 02:15:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

