

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081619\
 Data File : VN057331.D
 Acq On : 15 Aug 2019 15:03
 Operator : JC/SP
 Sample : K4264-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-3-C3-(3)

Quant Time: Aug 16 02:19:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1119785	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1566444	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1306762	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	332853	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	528362	43.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.80%	
35) Dibromofluoromethane	7.58	113	472673	46.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.06%	
50) Toluene-d8	10.09	98	1824922	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
62) 4-Bromofluorobenzene	12.40	95	480040	36.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.34%	
Target Compounds						
20) Methylene Chloride	4.54	84	16415	1.389	ug/l	88
43) Isopropyl Acetate	8.16	43	199277	8.687	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081619\
Data File : VN057331.D
Acq On : 15 Aug 2019 15:03
Operator : JC/SP
Sample : K4264-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-3-C3-(3)

Quant Time: Aug 16 02:19:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

