

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081619\
 Data File : VN057333.D
 Acq On : 15 Aug 2019 15:52
 Operator : JC/SP
 Sample : K4264-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-3-G1-(2)

Quant Time: Aug 16 02:23:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1152406	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1605005	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1361857	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	354934	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	538754	43.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.00%	
35) Dibromofluoromethane	7.58	113	481468	45.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.52%	
50) Toluene-d8	10.08	98	1898402	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	12.40	95	503064	36.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.98%	
Target Compounds						
20) Methylene Chloride	4.55	84	21589	1.775	ug/l	90
43) Isopropyl Acetate	8.15	43	207348	8.840	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081619\
Data File : VN057333.D
Acq On : 15 Aug 2019 15:52
Operator : JC/SP
Sample : K4264-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-3-G1-(2)

Quant Time: Aug 16 02:23:44 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

