

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081619\
 Data File : VN057336.D
 Acq On : 15 Aug 2019 17:05
 Operator : JC/SP
 Sample : K4264-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-3-G4-(2)

Quant Time: Aug 16 02:29:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1107966	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1552174	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1281327	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	317972	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	519546	43.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.26%	
35) Dibromofluoromethane	7.58	113	460130	45.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.44%	
50) Toluene-d8	10.09	98	1799539	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
62) 4-Bromofluorobenzene	12.40	95	467896	35.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.16%	
Target Compounds						
20) Methylene Chloride	4.54	84	16706	1.429	ug/l #	90
43) Isopropyl Acetate	8.16	43	185527	8.093	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081619\
Data File : VN057336.D
Acq On : 15 Aug 2019 17:05
Operator : JC/SP
Sample : K4264-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-3-G4-(2)

Quant Time: Aug 16 02:29:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

