

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081619\
 Data File : VN057342.D
 Acq On : 15 Aug 2019 19:31
 Operator : JC/SP
 Sample : K4264-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-3-C4-(2)

Quant Time: Aug 16 03:24:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1085453	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1527469	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1248496	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	311113	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	510628	43.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.54%	
35) Dibromofluoromethane	7.58	113	454662	45.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.82%	
50) Toluene-d8	10.08	98	1751252	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
62) 4-Bromofluorobenzene	12.40	95	449224	34.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.42%	
Target Compounds						
20) Methylene Chloride	4.54	84	22175	1.936	ug/l	95
43) Isopropyl Acetate	8.16	43	191116	8.525	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081619\
Data File : VN057342.D
Acq On : 15 Aug 2019 19:31
Operator : JC/SP
Sample : K4264-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-3-C4-(2)

Quant Time: Aug 16 03:24:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

