

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
 Data File : VN068108.D
 Acq On : 16 Aug 2021 11:46
 Operator : JC/MD
 Sample : VN0816WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0816WBL01

Quant Time: Aug 17 07:09:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	256486	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	487142	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	493123	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	163642	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	212705	58.580	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 117.160%	
35) Dibromofluoromethane	8.024	113	158977	55.998	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 112.000%	
50) Toluene-d8	10.443	98	664773	50.681	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.360%	
62) 4-Bromofluorobenzene	12.733	95	208053	44.041	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 88.080%	
Target Compounds						
95) Naphthalene	15.507	128	12245	4.803	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
Data File : VN068108.D
Acq On : 16 Aug 2021 11:46
Operator : JC/MD
Sample : VN0816WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0816WBL01

Quant Time: Aug 17 07:09:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

