

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
 Data File : VN068114.D
 Acq On : 16 Aug 2021 14:34
 Operator : JC/MD
 Sample : PB138399ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138399ZHE#03

Quant Time: Aug 17 07:12:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	252419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	510041	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	537169	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	185490	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	213041	59.618	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 119.240%			
35) Dibromofluoromethane	8.024	113	156273	52.574	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 105.140%			
50) Toluene-d8	10.446	98	693729	50.526	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 101.060%			
62) 4-Bromofluorobenzene	12.733	95	234295	46.994	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 93.980%			
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	42707	32.755	ug/l	100
18) Methyl Acetate	4.875	43	6980	2.129	ug/l #	67
20) Methylene Chloride	5.093	84	25815	8.338	ug/l #	87
43) Isopropyl Acetate	8.566	43	39821	5.382	ug/l #	89
95) Naphthalene	15.507	128	6230	3.875	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
Data File : VN068114.D
Acq On : 16 Aug 2021 14:34
Operator : JC/MD
Sample : PB138399ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138399ZHE#03

Quant Time: Aug 17 07:12:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

