

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
 Data File : VN068116.D
 Acq On : 16 Aug 2021 15:23
 Operator : JC/MD
 Sample : PB138399ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138399ZHE#05

Quant Time: Aug 17 07:13:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	231404	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	467976	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	497643	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	173963	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	198117	60.477	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 120.960%	
35) Dibromofluoromethane	8.024	113	144250	52.891	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.780%	
50) Toluene-d8	10.443	98	644847	51.137	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.280%	
62) 4-Bromofluorobenzene	12.734	95	216504	47.293	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.580%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	39431	32.988	ug/l	95
18) Methyl Acetate	4.865	43	6961	2.325	ug/l	91
20) Methylene Chloride	5.093	84	24267	8.557	ug/l	92
43) Isopropyl Acetate	8.566	43	38984	5.743	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
Data File : VN068116.D
Acq On : 16 Aug 2021 15:23
Operator : JC/MD
Sample : PB138399ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138399ZHE#05

Quant Time: Aug 17 07:13:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

