

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
 Data File : VN068117.D
 Acq On : 16 Aug 2021 15:47
 Operator : JC/MD
 Sample : PB138399ZHE#06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138399ZHE#06

Quant Time: Aug 17 07:13:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	224218	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	461940	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	478962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	163429	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	193818	61.061	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	122.120%
35) Dibromofluoromethane	8.024	113	140419	52.159	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.320%
50) Toluene-d8	10.443	98	631345	50.752	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.500%
62) 4-Bromofluorobenzene	12.733	95	201452	44.866	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.740%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	39154	33.807	ug/l	98
18) Methyl Acetate	4.862	43	9951	3.478	ug/l #	69
20) Methylene Chloride	5.087	84	24724	9.012	ug/l	88
43) Isopropyl Acetate	8.563	43	40795	6.088	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
Data File : VN068117.D
Acq On : 16 Aug 2021 15:47
Operator : JC/MD
Sample : PB138399ZHE#06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138399ZHE#06

Quant Time: Aug 17 07:13:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

