

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
 Data File : VN068122.D
 Acq On : 16 Aug 2021 17:50
 Operator : JC/MD
 Sample : PB138399ZHE#11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138399ZHE#11

Quant Time: Aug 17 07:15:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	233135	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	467461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	499217	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	169789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	195025	59.091	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 118.180%			
35) Dibromofluoromethane	8.024	113	145362	53.358	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 106.720%			
50) Toluene-d8	10.443	98	641224	50.923	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 101.840%			
62) 4-Bromofluorobenzene	12.733	95	211360	46.333	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 92.660%			
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	37238	30.922	ug/l	98
18) Methyl Acetate	4.870	43	11991	4.047	ug/l	96
20) Methylene Chloride	5.095	84	26992	9.477	ug/l	98
43) Isopropyl Acetate	8.566	43	35174	5.187	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081621\
Data File : VN068122.D
Acq On : 16 Aug 2021 17:50
Operator : JC/MD
Sample : PB138399ZHE#11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138399ZHE#11

Quant Time: Aug 17 07:15:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

